
Midmark M11 - v1.0.0
Total Cycles: 2419

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 11:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 180.4 F 3.6
4:00 191.1 F 7.7
6:00 210.9 F 12.2
8:00 223.4 F 5.8
10:00 234.0 F 10.7
12:00 234.1 F 6.9
14:00 243.4 F 12.9
16:00 241.7 F 11.2
18:00 254.1 F 17.6
20:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 29.1
2:00 271.6 F 28.7
3:00 272.3 F 29.1
4:00 271.5 F 28.6

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 223.9 F 0.0
10:00 238.7 F 0.0
15:00 233.5 F 0.0
20:00 227.3 F 0.0
25:00 225.3 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

10/12/2022 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____