
Midmark M11 - v1.0.0
Total Cycles: 1206

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2021 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 182.3 F 3.8
4:00 199.5 F 9.0
6:00 223.1 F 8.4
8:00 231.8 F 10.0
10:00 233.2 F 7.9
12:00 245.0 F 12.8
14:00 244.1 F 12.5
16:00 257.0 F 19.2
18:00 268.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.9
2:00 272.0 F 29.0
3:00 271.8 F 28.6
4:00 272.3 F 29.3

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 212.0 F 0.0
10:00 219.6 F 0.0
15:00 210.2 F 0.0
20:00 208.8 F 0.0
25:00 209.5 F 0.0
30:00 201.6 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

09/27/2021 11:11 AM
ACCEPTED: X
REJECTED:
Operator: U01