
Midmark M11 - v1.0.0
Total Cycles: 4354

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2024 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.8 F 0.0
2:00 152.3 F 2.4
4:00 159.3 F 5.5
6:00 205.8 F 9.4
8:00 214.8 F 9.5
10:00 223.4 F 6.4
12:00 231.3 F 10.7
14:00 230.5 F 7.4
16:00 242.3 F 12.1
18:00 240.1 F 10.2
20:00 252.1 F 16.1
22:00 262.6 F 22.2
24:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.5
1:00 272.7 F 28.7
2:00 272.6 F 28.9
3:00 271.6 F 28.8
4:00 271.6 F 28.4

Min 271.3 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 206.7 F 0.0
10:00 226.9 F 0.0
15:00 225.9 F 0.0
20:00 224.9 F 0.0
25:00 215.5 F 0.0
30:00 204.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 25:00
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

03/01/2024 05:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____