
Midmark M11 - v1.0.5
Total Cycles: 2341

Name: RITTER2
Sterilizer ID: V2313228
05/04/2023 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 180.7 F 4.0
4:00 194.5 F 8.8
6:00 215.5 F 8.5
8:00 227.8 F 10.1
10:00 231.7 F 7.9
12:00 242.1 F 12.0
14:00 245.2 F 13.4
16:00 258.3 F 20.5
18:00 269.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.3 F 29.4
2:00 272.2 F 29.8
3:00 271.7 F 29.1
4:00 272.5 F 29.8

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 214.5 F 0.0
10:00 228.4 F 0.0
15:00 233.1 F 0.0
20:00 229.6 F 0.0
25:00 222.8 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

05/04/2023 03:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____