
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001611
14:09 11/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:17
PRESSURIZING 7:11
STERILIZING 8:55
136.0 C 220kPa 8:55
136.4 C 221kPa 9:25
136.4 C 217kPa 9:55
136.4 C 222kPa 10:25
136.4 C 218kPa 10:55
136.4 C 222kPa 11:25
136.3 C 221kPa 11:55
136.1 C 220kPa 12:25
VENTING 12:25
AIR DRYING 13:32
CYCLE COMPLETE 22:27

Digital Signature #
4AC57E290672000C
