
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004708
15:21 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:58
STERILIZING 3:01
136.4 C 220kPa 3:01
136.4 C 225kPa 3:31
136.5 C 225kPa 4:01
136.4 C 225kPa 4:31
136.3 C 223kPa 5:01
136.2 C 221kPa 5:31
136.4 C 219kPa 6:01
136.2 C 225kPa 6:31
136.2 C 225kPa 6:31
VENTING 6:31
AIR DRYING 7:19
CYCLE COMPLETE 12:56

Digital Signature #
3B47A1DB85C436CC
