
Midmark M11 - v1.0.5
Total Cycles: 744

Name: RITTER2
Sterilizer ID: V2313228
02/01/2022 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.4 F 0.0
2:00 180.4 F 3.6
4:00 202.9 F 9.3
6:00 220.2 F 7.0
8:00 233.8 F 10.6
10:00 235.6 F 9.0
12:00 238.7 F 9.6
14:00 251.5 F 16.1
16:00 264.1 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 271.8 F 29.1
2:00 272.4 F 29.0
3:00 271.4 F 28.4
4:00 271.4 F 28.9

Min 271.0 F 28.3
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 231.7 F 0.0
10:00 209.7 F 0.0
15:00 195.9 F 0.0
20:00 192.0 F 0.0
25:00 195.6 F 0.0
30:00 201.9 F 0.0

Drying Complete

FILLING: 1:51
HEATING: 17:47
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:56:43
CYCLE COMPLETE

02/01/2022 12:42 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____