
Midmark M11 - v1.0.0
Total Cycles: 5492

Name: RITTER 1
Sterilizer ID: V2178198
04/09/2025 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 112.3 F 0.9
4:00 123.3 F 3.3
6:00 142.5 F 6.3
8:00 170.2 F 9.6
10:00 192.8 F 13.0
12:00 207.6 F 4.6
14:00 215.0 F 8.5
16:00 227.4 F 12.7
18:00 230.9 F 8.1
20:00 242.3 F 13.0
22:00 239.7 F 10.8
24:00 251.2 F 16.5
26:00 261.6 F 22.4
28:00 271.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.8 F 29.2
2:00 272.6 F 29.6
3:00 272.2 F 29.1
4:00 271.3 F 28.7

Min 271.3 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 188.4 F 0.0
10:00 181.5 F 0.0
15:00 190.7 F 0.0
20:00 194.9 F 0.0
25:00 203.3 F 0.0
30:00 209.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:32
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:07:14
CYCLE COMPLETE

04/09/2025 05:21 PM
ACCEPTED: X
REJECTED:
Operator: U01