
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000534
16:36 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:28
STERILIZING 4:46
136.0 C 229kPa 4:46
136.3 C 231kPa 5:16
136.1 C 230kPa 5:46
136.3 C 231kPa 6:16
136.0 C 230kPa 6:46
136.4 C 233kPa 7:16
136.2 C 232kPa 7:46
136.3 C 232kPa 8:16
136.3 C 232kPa 8:16
VENTING 8:16
AIR DRYING 9:12
CYCLE COMPLETE 24:12

Digital Signature #
C103B66B75E19ACF
