
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001259
10:59 24/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:35
STERILIZING 3:38
136.4 C 222kPa 3:38
136.1 C 222kPa 4:08
136.2 C 222kPa 4:38
136.3 C 224kPa 5:08
136.4 C 224kPa 5:38
136.2 C 223kPa 6:08
136.4 C 224kPa 6:38
136.2 C 222kPa 7:08
136.2 C 222kPa 7:08
VENTING 7:08
AIR DRYING 8:01
CYCLE COMPLETE 8:26

Digital Signature #
35F8BAB68741AA51
