

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2539

Name: RITTER 1  
Sterilizer ID: V2178198  
11/10/2022 11:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
2:00 194.4 F 4.0  
4:00 195.3 F 8.6  
6:00 215.8 F 10.8  
8:00 227.6 F 8.6  
10:00 238.1 F 11.1  
12:00 240.0 F 11.4  
14:00 238.3 F 9.7  
16:00 251.9 F 16.3  
18:00 263.5 F 23.2  
20:00 271.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.2  
2:00 271.8 F 28.5  
3:00 272.0 F 29.1  
4:00 272.3 F 29.0  
  
Min 271.3 F 28.3  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 227.5 F 0.0  
10:00 231.5 F 0.0  
15:00 233.0 F 0.0  
20:00 225.8 F 0.0  
25:00 223.8 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

11/10/2022 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_