

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000834  
17:36 22/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:36  
STERILIZING 4:56  
136.0 C 220kPa 4:56  
136.4 C 233kPa 5:26  
136.2 C 231kPa 5:56  
136.4 C 233kPa 6:26  
136.2 C 232kPa 6:56  
136.4 C 233kPa 7:26  
136.4 C 233kPa 7:56  
136.4 C 232kPa 8:26  
136.4 C 232kPa 8:26  
VENTING 8:26  
AIR DRYING 9:21  
CYCLE COMPLETE 13:32

-----  
Digital Signature #  
1E6088E0059734B0  
-----