
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005971
8:06 14/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:36
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 221kPa 4:08
136.4 C 223kPa 4:38
136.4 C 222kPa 5:08
136.2 C 221kPa 5:38
136.4 C 222kPa 6:08
136.1 C 221kPa 6:38
136.3 C 222kPa 7:08
136.3 C 222kPa 7:08
VENTING 7:08
AIR DRYING 7:58
CYCLE COMPLETE 9:31

Digital Signature #
91784CB0457A9CD5
