
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002097
9:51 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:23
STERILIZING 3:43
136.4 C 226kPa 3:43
136.4 C 226kPa 4:13
136.4 C 225kPa 4:43
136.3 C 224kPa 5:13
136.2 C 223kPa 5:43
136.4 C 222kPa 6:13
136.4 C 225kPa 6:43
136.4 C 225kPa 7:13
136.4 C 225kPa 7:13
VENTING 7:13
AIR DRYING 8:07
CYCLE COMPLETE 31:56

Digital Signature #
9C4209ED4A49AB4E
