
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000274
8:40 23/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:24
STERILIZING 4:28
136.0 C 221kPa 4:28
136.4 C 225kPa 4:58
136.4 C 224kPa 5:28
136.3 C 223kPa 5:58
136.2 C 218kPa 6:28
136.4 C 221kPa 6:58
136.2 C 224kPa 7:28
136.4 C 223kPa 7:58
136.4 C 223kPa 7:58
VENTING 7:58
AIR DRYING 8:48
CYCLE COMPLETE 8:57

Digital Signature #
B1ADA16EE3751209
