

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 000053  
8:34 13/03/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:52  
PRESSURIZING 2:58  
STERILIZING 4:24  
136.4 C 224kPa 4:24  
136.4 C 224kPa 4:55  
136.3 C 224kPa 5:25  
136.1 C 222kPa 5:55  
136.2 C 223kPa 6:25  
136.2 C 223kPa 6:55  
136.4 C 224kPa 7:25  
136.4 C 225kPa 7:55  
136.4 C 225kPa 7:55  
VENTING 7:55  
AIR DRYING 8:46  
CYCLE COMPLETE 23:46

-----  
Digital Signature #  
D11665A08008CBAC  
-----