
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003499
8:43 21/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:34
STERILIZING 3:28
136.0 C 220kPa 3:28
136.2 C 220kPa 3:58
136.3 C 218kPa 4:28
136.1 C 219kPa 4:58
136.4 C 223kPa 5:28
136.1 C 221kPa 5:58
136.0 C 217kPa 6:28
136.1 C 219kPa 6:58
136.1 C 219kPa 6:58
VENTING 6:58
AIR DRYING 7:56
CYCLE COMPLETE 8:05

Digital Signature #
29D3F44BE607EC42
