

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3530

Name: RITTER2  
Sterilizer ID: V2313228  
04/15/2024 04:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.9 F 0.0  
2:00 130.9 F 1.7  
4:00 153.2 F 4.8  
6:00 185.1 F 8.7  
8:00 201.5 F 12.9  
10:00 215.1 F 6.0  
12:00 227.8 F 10.6  
14:00 229.4 F 7.7  
16:00 240.7 F 12.9  
18:00 242.1 F 11.7  
20:00 254.4 F 18.0  
22:00 264.7 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.7 F 29.9  
2:00 271.3 F 29.2  
3:00 272.7 F 30.0  
4:00 271.7 F 29.4

Min 271.2 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 215.2 F 0.0  
10:00 213.6 F 0.0  
15:00 191.1 F 0.0  
20:00 177.0 F 0.0  
25:00 175.3 F 0.0  
30:00 175.7 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 23:55  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

04/15/2024 05:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_