
Midmark M11 - v1.0.0
Total Cycles: 1750

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 07:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 73.5 F 0.0
4:00 81.8 F 0.1
6:00 114.0 F 0.7
8:00 145.9 F 2.3
10:00 171.2 F 4.6
12:00 190.6 F 7.7
14:00 205.2 F 12.1
16:00 214.8 F 5.1
18:00 226.7 F 9.2
20:00 237.1 F 12.1
22:00 237.0 F 10.5
24:00 233.9 F 7.8
26:00 248.9 F 14.9
28:00 260.6 F 21.5
30:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.3
2:00 271.8 F 29.0
3:00 272.5 F 29.2
4:00 271.8 F 29.0

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 210.9 F 0.0
10:00 191.5 F 0.0
15:00 193.3 F 0.0
20:00 222.9 F 0.0
25:00 228.3 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:35
CYCLE COMPLETE

03/28/2022 08:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____