
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000093
13:52 13/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:24
STERILIZING 4:43
136.4 C 221kPa 4:43
136.3 C 224kPa 5:13
136.4 C 223kPa 5:43
136.2 C 223kPa 6:13
136.3 C 225kPa 6:43
136.4 C 221kPa 7:13
136.1 C 222kPa 7:43
136.3 C 223kPa 8:13
136.3 C 223kPa 8:13
VENTING 8:13
AIR DRYING 9:07
CYCLE COMPLETE 9:18

Digital Signature #
88E82EC89B9E4A43
