
Midmark M11 - v1.0.5
Total Cycles: 22

Name: RITTER2
Sterilizer ID: V2313228
04/27/2021 06:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 74.6 F 0.3
4:00 77.5 F 1.2
6:00 94.8 F 3.5
8:00 131.7 F 6.5
10:00 158.7 F 10.1
12:00 185.0 F 5.5
14:00 206.6 F 6.0
16:00 220.3 F 10.2
18:00 229.1 F 8.2
20:00 236.8 F 12.0
22:00 238.4 F 10.7
24:00 251.4 F 17.2
26:00 262.9 F 24.0
28:00 273.4 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 271.6 F 29.8
2:00 272.7 F 30.4
3:00 272.2 F 29.7
4:00 271.5 F 29.4

Min 271.3 F 29.3
Max 273.4 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 229.1 F 0.0
10:00 225.4 F 0.0
15:00 191.5 F 0.0
20:00 184.3 F 0.0
25:00 184.2 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 28:09
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:06:14
CYCLE COMPLETE

04/27/2021 07:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____