
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001686
9:56 07/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:43
136.5 C 223kPa 4:43
136.3 C 223kPa 5:13
136.4 C 224kPa 5:43
136.2 C 221kPa 6:13
136.5 C 225kPa 6:43
136.1 C 221kPa 7:13
136.2 C 222kPa 7:43
136.4 C 224kPa 8:13
136.4 C 224kPa 8:13
VENTING 8:14
AIR DRYING 9:09
CYCLE COMPLETE 9:11

Digital Signature #
95AEF7C2CE643F65
