
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000156
9:42 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:43
136.4 C 223kPa 5:43
136.2 C 224kPa 6:13
136.2 C 224kPa 6:43
136.4 C 223kPa 7:13
136.3 C 224kPa 7:43
136.2 C 223kPa 8:13
136.2 C 223kPa 8:43
136.3 C 224kPa 9:13
136.3 C 224kPa 9:13
VENTING 9:13
AIR DRYING 10:16
CYCLE COMPLETE 10:19

Digital Signature #
8CAE4A7B165530E3
