
Midmark M11 - v1.0.0
Total Cycles: 1822

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2022 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.3 F 0.0
2:00 137.9 F 0.2
4:00 153.8 F 3.3
6:00 182.4 F 7.4
8:00 206.0 F 12.5
10:00 221.4 F 5.9
12:00 232.3 F 10.7
14:00 232.2 F 7.0
16:00 244.5 F 13.0
18:00 242.5 F 11.5
20:00 255.4 F 18.3
22:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.8
2:00 271.2 F 28.4
3:00 272.4 F 28.8
4:00 271.1 F 28.2

Min 271.1 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 217.1 F 0.0
10:00 204.0 F 0.0
15:00 211.9 F 0.0
20:00 219.4 F 0.0
25:00 238.8 F 0.0
30:00 238.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

04/13/2022 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____