
Midmark M11 - v1.0.0
Total Cycles: 4104

Name: RITTER 1
Sterilizer ID: V2178198
12/11/2023 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 138.3 F 1.9
4:00 160.2 F 4.6
6:00 203.5 F 8.2
8:00 210.8 F 12.4
10:00 217.8 F 5.4
12:00 229.1 F 9.6
14:00 237.7 F 10.6
16:00 238.5 F 10.9
18:00 236.2 F 8.7
20:00 248.4 F 14.9
22:00 260.0 F 21.0
24:00 270.1 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.2 F 29.0
2:00 272.7 F 29.4
3:00 271.7 F 29.3
4:00 271.9 F 29.0

Min 271.4 F 28.9
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.1 F 0.0
5:00 226.1 F 0.0
10:00 216.6 F 0.0
15:00 230.7 F 0.0
20:00 229.2 F 0.0
25:00 232.6 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:34
CYCLE COMPLETE

12/11/2023 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____