
Midmark M11 - v1.0.5
Total Cycles: 3620

Name: RITTER2
Sterilizer ID: V2313228
05/15/2024 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 181.3 F 2.5
4:00 184.8 F 7.0
6:00 198.5 F 12.2
8:00 220.1 F 7.3
10:00 231.2 F 12.8
12:00 235.4 F 10.4
14:00 236.0 F 9.1
16:00 249.7 F 16.0
18:00 261.7 F 22.8
20:00 272.4 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 30.2
1:00 272.2 F 30.3
2:00 272.3 F 30.4
3:00 272.2 F 30.3
4:00 272.2 F 30.3

Min 272.2 F 30.1
Max 273.1 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 231.9 F 0.0
10:00 225.5 F 0.0
15:00 225.8 F 0.0
20:00 222.9 F 0.0
25:00 232.7 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

05/15/2024 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____