
Midmark M11 - v1.0.0
Total Cycles: 3465

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.3 F 0.0
2:00 198.8 F 3.6
4:00 216.0 F 8.5
6:00 224.8 F 7.8
8:00 234.4 F 9.9
10:00 235.9 F 8.2
12:00 245.2 F 12.8
14:00 244.9 F 12.7
16:00 257.9 F 19.6
18:00 269.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.1
2:00 272.1 F 28.5
3:00 271.5 F 28.4
4:00 272.7 F 29.2

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 225.8 F 0.0
10:00 235.2 F 0.0
15:00 223.8 F 0.0
20:00 224.0 F 0.0
25:00 223.2 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

06/29/2023 11:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____