

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3082

Name: RITTER2  
Sterilizer ID: V2313228  
11/03/2023 11:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.2 F 0.0  
2:00 133.7 F 1.3  
4:00 159.7 F 4.6  
6:00 188.4 F 9.3  
8:00 206.2 F 7.4  
10:00 220.4 F 8.0  
12:00 233.9 F 13.3  
14:00 237.4 F 10.6  
16:00 236.4 F 9.2  
18:00 251.3 F 16.4  
20:00 263.7 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 273.0 F 29.4  
2:00 271.4 F 28.7  
3:00 272.7 F 29.6  
4:00 271.8 F 28.8  
  
Min 271.4 F 28.6  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.8 F 0.0  
5:00 203.7 F 0.0  
10:00 207.1 F 0.0  
15:00 205.0 F 0.0  
20:00 201.9 F 0.0  
25:00 199.1 F 0.0  
30:00 198.3 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 21:57  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

11/03/2023 12:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_