
Midmark M11 - v1.0.5
Total Cycles: 3892

Name: RITTER2
Sterilizer ID: V2313228
08/26/2024 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.3 F 0.0
2:00 91.7 F 0.0
4:00 102.2 F 0.0
6:00 129.4 F 0.9
8:00 162.2 F 3.6
10:00 185.9 F 6.6
12:00 203.4 F 10.6
14:00 213.2 F 7.1
16:00 220.6 F 7.4
18:00 231.9 F 11.7
20:00 232.2 F 8.4
22:00 242.7 F 12.4
24:00 243.7 F 12.6
26:00 255.0 F 18.5
28:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.7
2:00 271.6 F 29.0
3:00 271.9 F 29.8
4:00 272.4 F 29.5

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 177.7 F 0.0
10:00 192.5 F 0.0
15:00 191.3 F 0.0
20:00 190.4 F 0.0
25:00 191.1 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:08:32
CYCLE COMPLETE

08/26/2024 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____