
Midmark M11 - v1.0.0
Total Cycles: 5464

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2025 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 119.6 F 1.5
4:00 141.2 F 4.3
6:00 168.4 F 7.8
8:00 187.5 F 11.7
10:00 205.7 F 3.2
12:00 214.8 F 7.3
14:00 226.8 F 11.2
16:00 228.3 F 6.0
18:00 239.5 F 11.4
20:00 235.0 F 8.7
22:00 247.9 F 14.6
24:00 258.8 F 20.5
26:00 268.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.2 F 29.4
2:00 271.5 F 28.6
3:00 271.9 F 29.5
4:00 272.0 F 29.1

Min 271.1 F 28.6
Max 272.3 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 202.2 F 0.0
10:00 203.8 F 0.0
15:00 201.5 F 0.0
20:00 209.9 F 0.0
25:00 219.2 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:11
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:41
CYCLE COMPLETE

03/31/2025 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____