
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000071
8:53 31/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:40
STERILIZING 4:59
136.0 C 220kPa 4:59
136.4 C 225kPa 5:29
136.1 C 222kPa 5:59
136.3 C 224kPa 6:29
136.3 C 223kPa 6:59
136.2 C 223kPa 7:29
136.3 C 222kPa 7:59
136.1 C 222kPa 8:29
136.1 C 222kPa 8:29
VENTING 8:30
AIR DRYING 9:23
CYCLE COMPLETE 12:52

Digital Signature #
DF515291618F22C5
