

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3389

Name: RITTER2  
Sterilizer ID: V2313228  
02/16/2024 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 69.5 F 0.0  
4:00 77.0 F 0.0  
6:00 106.7 F 1.1  
8:00 141.1 F 3.6  
10:00 167.4 F 6.4  
12:00 183.8 F 10.3  
14:00 201.7 F 6.7  
16:00 216.6 F 6.2  
18:00 229.1 F 10.0  
20:00 234.1 F 10.2  
22:00 238.7 F 10.9  
24:00 234.4 F 8.5  
26:00 248.2 F 14.7  
28:00 259.2 F 20.8  
30:00 269.2 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.9  
1:00 272.2 F 29.6  
2:00 271.4 F 28.7  
3:00 272.6 F 29.5  
4:00 271.2 F 28.8

Min 271.2 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 239.3 F 0.0  
10:00 225.0 F 0.0  
15:00 197.9 F 0.0  
20:00 184.9 F 0.0  
25:00 180.8 F 0.0  
30:00 175.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:56  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:57  
CYCLE COMPLETE

02/16/2024 10:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_