
Midmark M11 - v1.0.0
Total Cycles: 815

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 163.7 F 2.9
4:00 184.5 F 7.5
6:00 215.3 F 13.2
8:00 223.2 F 8.6
10:00 234.9 F 9.7
12:00 240.1 F 11.9
14:00 239.1 F 10.2
16:00 253.5 F 17.4
18:00 265.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.1 F 29.0
2:00 271.6 F 29.0
3:00 272.0 F 28.9
4:00 271.8 F 29.1

Min 271.1 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 222.2 F 0.0
10:00 243.5 F 0.0
15:00 225.4 F 0.0
20:00 224.6 F 0.0
25:00 225.0 F 0.0
30:00 216.2 F 0.0

Drying Complete

FILLING: 1:32
HEATING: 19:39
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/27/2021 01:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____