
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001485
16:08 07/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:20
136.0 C 220kPa 4:20
136.1 C 223kPa 4:50
136.4 C 225kPa 5:20
136.3 C 223kPa 5:50
136.2 C 223kPa 6:20
136.1 C 222kPa 6:50
136.2 C 221kPa 7:20
136.2 C 222kPa 7:50
136.2 C 222kPa 7:50
VENTING 7:50
AIR DRYING 8:44
CYCLE COMPLETE 8:53

Digital Signature #
F0257CD49ECA2BCF
