
Midmark M11 - v1.0.0
Total Cycles: 4822

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2024 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.1 F 0.0
2:00 102.5 F 0.1
4:00 128.7 F 0.7
6:00 149.2 F 1.3
8:00 190.0 F 3.4
10:00 212.3 F 6.9
12:00 223.0 F 10.6
14:00 228.5 F 6.6
16:00 231.2 F 8.7
18:00 241.4 F 13.3
20:00 239.5 F 10.3
22:00 236.5 F 8.4
24:00 248.4 F 14.2
26:00 259.2 F 20.2
28:00 268.9 F 26.3
30:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 271.6 F 28.6
2:00 272.5 F 29.2
3:00 272.7 F 28.9
4:00 271.6 F 28.5

Min 271.2 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 259.5 F 0.0
10:00 214.4 F 0.0
15:00 217.1 F 0.0
20:00 225.2 F 0.0
25:00 230.8 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

08/15/2024 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____