
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004941
11:25 10/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:07
136.4 C 222kPa 4:07
136.1 C 220kPa 4:37
136.3 C 224kPa 5:07
136.4 C 223kPa 5:37
136.2 C 223kPa 6:07
136.3 C 224kPa 6:37
136.4 C 218kPa 7:07
136.4 C 223kPa 7:37
136.4 C 223kPa 7:37
VENTING 7:37
AIR DRYING 8:27
CYCLE COMPLETE 11:21

Digital Signature #
14D4615F023774C9
