
Midmark M11 - v1.0.0
Total Cycles: 2421

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 202.7 F 4.3
4:00 225.3 F 9.1
6:00 219.2 F 8.2
8:00 228.8 F 8.6
10:00 237.2 F 11.3
12:00 240.5 F 11.4
14:00 238.6 F 9.7
16:00 252.2 F 16.5
18:00 263.8 F 23.3
20:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.0
2:00 271.4 F 28.4
3:00 272.3 F 29.1
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 233.2 F 0.0
10:00 242.1 F 0.0
15:00 237.1 F 0.0
20:00 226.9 F 0.0
25:00 222.9 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

10/12/2022 02:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____