
Midmark M11 - v1.0.0
Total Cycles: 1609

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.9 F 0.0
2:00 140.2 F 2.4
4:00 164.6 F 6.1
6:00 202.7 F 10.9
8:00 214.1 F 4.4
10:00 225.7 F 9.3
12:00 235.0 F 9.9
14:00 240.0 F 11.4
16:00 238.0 F 9.6
18:00 251.8 F 16.3
20:00 263.6 F 23.1
22:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.2
4:00 272.1 F 28.6

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 214.4 F 0.0
10:00 212.5 F 0.0
15:00 219.7 F 0.0
20:00 249.5 F 0.0
25:00 229.5 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

02/08/2022 03:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____