
Midmark M11 - v1.0.0
Total Cycles: 3067

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 146.5 F 0.5
4:00 163.3 F 2.5
6:00 188.1 F 6.0
8:00 206.7 F 10.8
10:00 219.3 F 5.0
12:00 230.4 F 10.1
14:00 234.3 F 8.2
16:00 245.1 F 12.8
18:00 242.2 F 11.2
20:00 255.1 F 17.8
22:00 266.4 F 24.6
24:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.7 F 28.6
2:00 271.7 F 28.6
3:00 271.7 F 28.5
4:00 271.9 F 28.6

Min 271.5 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 222.5 F 0.0
10:00 232.0 F 0.0
15:00 228.5 F 0.0
20:00 230.1 F 0.0
25:00 229.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:10
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

04/03/2023 01:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____