

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001031  
12:08 08/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:57  
STERILIZING 4:06  
136.0 C 220kPa 4:06  
136.2 C 223kPa 4:36  
136.0 C 222kPa 5:06  
136.4 C 226kPa 5:36  
136.4 C 223kPa 6:06  
136.3 C 225kPa 6:36  
136.3 C 224kPa 7:06  
136.4 C 225kPa 7:36  
136.4 C 225kPa 7:36  
VENTING 7:36  
AIR DRYING 8:29  
CYCLE COMPLETE 23:30  
-----

Digital Signature #  
10C4E623F64F4BCC  
-----