
Midmark M11 - v1.0.5
Total Cycles: 273

Name: RITTER2
Sterilizer ID: V2313228
08/09/2021 03:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.5 F 0.0
2:00 115.5 F 1.0
4:00 123.4 F 3.4
6:00 151.4 F 6.9
8:00 178.9 F 10.9
10:00 199.6 F 4.0
12:00 216.0 F 8.2
14:00 229.5 F 12.8
16:00 231.5 F 10.1
18:00 234.8 F 9.0
20:00 245.9 F 14.4
22:00 258.3 F 20.6
24:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.5 F 29.2
2:00 272.3 F 29.7
3:00 272.1 F 29.4
4:00 271.7 F 29.2

Min 270.8 F 28.7
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 235.8 F 0.0
10:00 222.9 F 0.0
15:00 226.0 F 0.0
20:00 221.0 F 0.0
25:00 217.0 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

08/09/2021 04:54 PM
ACCEPTED: X
REJECTED:
Operator: U01