
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001295
8:15 01/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:54
STERILIZING 5:06
136.4 C 221kPa 5:06
136.3 C 223kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:36
136.3 C 221kPa 7:06
136.4 C 223kPa 7:36
136.4 C 224kPa 8:06
136.4 C 223kPa 8:36
136.4 C 223kPa 8:36
VENTING 8:37
AIR DRYING 9:39
CYCLE COMPLETE 9:43

Digital Signature #
C9E6C04000D59F4E
