
Midmark M11 - v1.0.0
Total Cycles: 1727

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 184.7 F 3.6
4:00 201.4 F 8.2
6:00 226.6 F 8.6
8:00 232.5 F 9.6
10:00 235.8 F 8.8
12:00 243.8 F 13.1
14:00 242.5 F 11.8
16:00 255.6 F 18.7
18:00 267.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 29.2
2:00 271.6 F 28.6
3:00 272.7 F 29.3
4:00 271.8 F 28.6

Min 271.5 F 28.5
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.1 F 0.0
5:00 231.9 F 0.0
10:00 229.8 F 0.0
15:00 228.5 F 0.0
20:00 228.2 F 0.0
25:00 230.0 F 0.0
30:00 227.7 F 0.0

Drying Complete

FILLING: 2:18
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

03/16/2022 05:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____