
Midmark M11 - v1.0.5
Total Cycles: 710

Name: RITTER2
Sterilizer ID: V2313228
01/20/2022 07:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 76.5 F 0.2
4:00 82.5 F 1.3
6:00 113.8 F 3.6
8:00 148.9 F 6.8
10:00 170.0 F 10.7
12:00 202.0 F 4.1
14:00 212.6 F 7.5
16:00 223.9 F 11.7
18:00 227.2 F 7.5
20:00 239.8 F 12.7
22:00 239.2 F 10.8
24:00 251.8 F 17.1
26:00 262.7 F 23.4
28:00 272.7 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.7 F 30.3
2:00 272.2 F 30.4
3:00 272.5 F 30.4
4:00 272.4 F 30.4

Min 270.9 F 29.1
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 199.0 F 0.0
10:00 195.7 F 0.0
15:00 190.9 F 0.0
20:00 188.8 F 0.0
25:00 188.0 F 0.0
30:00 187.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:24
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:06:18
CYCLE COMPLETE

01/20/2022 08:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____