
Midmark M11 - v1.0.0
Total Cycles: 3400

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 84.9 F 0.0
4:00 118.2 F 0.0
6:00 130.4 F 0.0
8:00 166.0 F 1.0
10:00 197.0 F 4.9
12:00 206.3 F 8.8
14:00 218.9 F 12.8
16:00 224.2 F 6.2
18:00 234.4 F 10.1
20:00 238.3 F 9.0
22:00 242.2 F 11.6
24:00 238.5 F 9.6
26:00 252.2 F 16.2
28:00 264.0 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 271.8 F 28.4
2:00 272.3 F 28.7
3:00 272.6 F 28.9
4:00 272.7 F 28.8

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 233.1 F 0.0
Dry Cycle Aborted

FILLING: 1:01
HEATING: 29:51
STERILIZING: 4:00
VENTING: 2:51
DRYING: 1:11
TOTAL CYCLE: 00:38:55

06/16/2023 09:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____