
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000421
11:00 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:49
STERILIZING 4:53
136.0 C 220kPa 4:53
136.1 C 222kPa 5:23
136.4 C 223kPa 5:53
136.4 C 223kPa 6:23
136.3 C 223kPa 6:53
136.3 C 223kPa 7:23
136.2 C 222kPa 7:53
136.4 C 224kPa 8:23
136.4 C 224kPa 8:23
VENTING 8:23
AIR DRYING 9:15
CYCLE COMPLETE 9:32

Digital Signature #
9F7C9935A8675057
