
Midmark M11 - v1.0.5
Total Cycles: 681

Name: RITTER2
Sterilizer ID: V2313228
01/11/2022 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 185.9 F 3.8
4:00 205.7 F 9.2
6:00 224.6 F 7.7
8:00 234.1 F 11.2
10:00 235.8 F 9.4
12:00 237.7 F 9.3
14:00 250.7 F 15.8
16:00 263.0 F 23.0
18:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.3
2:00 271.4 F 28.9
3:00 272.3 F 29.3
4:00 271.5 F 28.7

Min 271.4 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 236.1 F 0.0
10:00 223.2 F 0.0
15:00 224.3 F 0.0
20:00 224.1 F 0.0
25:00 228.6 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 3:46
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

01/11/2022 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____