
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000825
13:36 30/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:37
PRESSURIZING 6:31
STERILIZING 8:17
136.4 C 223kPa 8:17
136.4 C 223kPa 8:47
136.4 C 223kPa 9:17
135.6 C 219kPa 9:47
136.3 C 222kPa 10:17
136.0 C 221kPa 10:47
136.3 C 221kPa 11:17
136.2 C 222kPa 11:47
136.2 C 222kPa 11:47
VENTING 11:47
AIR DRYING 12:48
CYCLE COMPLETE 13:24

Digital Signature #
09F2402042A7BDB5
