
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000239
9:08 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:00
PRESSURIZING 4:01
STERILIZING 5:33
136.4 C 233kPa 5:33
136.2 C 231kPa 6:03
136.4 C 233kPa 6:33
136.4 C 233kPa 7:03
136.4 C 233kPa 7:33
136.4 C 233kPa 8:03
136.2 C 230kPa 8:33
136.1 C 231kPa 9:03
136.1 C 231kPa 9:03
VENTING 9:03
AIR DRYING 10:06
CYCLE COMPLETE 16:17

Digital Signature #
236061090DC192D9
