
Midmark M11 - v1.0.5
Total Cycles: 3964

Name: RITTER2
Sterilizer ID: V2313228
09/20/2024 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 159.4 F 2.6
4:00 172.2 F 6.3
6:00 195.5 F 10.7
8:00 217.7 F 5.7
10:00 227.4 F 10.0
12:00 231.5 F 8.5
14:00 240.6 F 12.6
16:00 240.5 F 11.3
18:00 252.9 F 17.7
20:00 263.9 F 24.3
22:00 271.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 29.9
2:00 271.5 F 29.7
3:00 272.2 F 29.8
4:00 271.5 F 29.7

Min 271.3 F 29.3
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 206.2 F 0.0
10:00 230.7 F 0.0
15:00 227.6 F 0.0
20:00 224.9 F 0.0
25:00 223.7 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

09/20/2024 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____